





Medical protective equipment)——Purchase list

Date: 2020/02/11; Valid date: 2020/06/30

Customer: Company Name: CHINA HOLDINGS CO., LTD.; Detail address: Beijing Grain Building, 16 East Third Ring Road, Chaoyang District, Beijing; Zip Code: 100022; Contact Name: Guo Shaokui; Phone number: +86-10-80303697; Fax Number: +86-10-80303697; Mobile number: +86-13699225904; e-mail: 13699225904@qq.com; Chat: Messenger/Online/WeChat/Skype/WhatsApp; Business Philosophy: Integrity, Innovation, Beyond, Win-Win; Service Concept: Quality First, Scientific management, thoughtful service, customer satisfaction.

Manufacturer: Factory logo, company name, detailed address, zip code, contact name, job name, phone number, fax number, mobile phone number, email address, chat mode, company URL.

(No.)	Product's Name	(Photo)	(Description)	(Quality Terms)	Unit	(MOQ)	CIF China Port USD/PCS (USD=7.92RMB)	(Subtotal) (USD)	备注
1	Medical ordinary mask		Scope of application: the protection of pathogenic microorganisms is not accurate, which can be used for one-time health care in general environment, or for the barrier and protection of particles other than pathogenic microorganisms, such as flower powder. Generally, there is no requirement for the filtration efficiency of particles and bacteria, or the filtration efficiency of particles and bacteria is lower than that of medical surgical masks and medical protective masks.	China: YY / T 0969-2013 disposable medical masks, filtration efficiency requirements (0.3um), gas flow requirements (28.3l / min), 95% BFE;	PCS	1 million			Tax code column 6307.9000
2	surgical mask		Scope of application: it is applicable to the basic protection of medical personnel and related personnel, as well as the protection of tissue blood, body fluid and spatter transmission in the process of invasive operation; Important technical indicators include filtration efficiency, bacterial filtration efficiency and respiratory resistance: (1) filtration efficiency: under the condition of air flow (30S5 ± 2)L/min, the filtration efficiency of sodium chloride aerosol with median diameter (0.24 ± 0.06) μ m in aerodynamics is not less than 30%; (2) bacterial filtration efficiency: under the specified conditions, the Golden Grape ball with average particle diameter (3 ± 0.3) μ m is used The filtration efficiency of bacterial aerosol is not less than 95%; (3) respiratory resistance: under the condition of filtration efficiency flow, the suction resistance is not more than 49pa, and the expiratory resistance is not more than 29.4pa.	China: YY 0469-2011 Technical Requirements for Medical Surgical Masks, Filtration Efficiency Requirements (0.3um), Gas Flow Requirements (28.3L/min), 95% BFE (PFE-30%); Overseas Standard: ASTM F2100-II Standard, level2-level3;	PCS	1 million			Tax code column 6307.9000





Medical protective equipment)——Purchase list

Date: 2020/02/11; Valid date: 2020/06/30

Customer: Company Name: CHINA HOLDINGS CO., LTD.; Detail address: Beijing Grain Building, 16 East Third Ring Road, Chaoyang District, Beijing; Zip Code: 100022; Contact Name: Guo Shaokui; Phone number: +86-10-80303697; Fax Number: +86-10-80303697; Mobile number: +86-13699225904; e-mail: 13699225904@qq.com; Chat: Messenger/Online/WeChat/Skype/WhatsApp; Business Philosophy: Integrity, Innovation, Beyond, Win-Win; Service Concept: Quality First, Scientific management, thoughtful service, customer satisfaction.

Manufacturer: Factory logo, company name, detailed address, zip code, contact name, job name, phone number, fax number, mobile phone number, email address, chat mode, company URL.

(No.)	Product's Name	(Photo)	(Description)	(Quality Terms)	Unit	(MOQ)	CIF China Port USD/PCS (USD=7.92RMB)	(Subtotal) (USD)	备注
3	Medical protective mask		Scope of application: it is applicable to the protection of medical staff and relevant staff for respiratory infectious diseases transmitted by air, with high protection level. The main technical indexes include filtration efficiency and air flow resistance of non oil particles: (1) filtration efficiency: under the condition of air flow (85±2)L/min, the filtration efficiency of sodium chloride aerosol with median aerodynamic diameter (0.24 ± 0.06) μ m is not less than 95%, that is to say, it meets the grade of N95 (or FFP2) and above. (2) Suction resistance: under the above flow conditions, the suction resistance shall not exceed 343.2pa (35mmH ₂ O).	China: GB 19063-2010 Technical Requirements for Medical Protective Masks; United States: ASTM F2100-2011; Airflow requirements are 85L/min; 95% LOW (BFE), 98% moderate, High (BFE, PFE); EU: EN14683-2014; Airflow requirement ≥28.3L/min; 95% TYPE LA: 98% TYPE1L, TYPE1L R; Australia: AS1716; Japan: JS-T-8151	PCS	1 million			Tax code column 6307.9000
4	Medical eyepieces		Also known as protective glasses, it protects against liquid spills, with flat-light lenses and frames covered.	Level 2 or above medical eyepieces	PCS	1 million			Tax code number 9004.9090